



PRODUCT MANUAL FOR

Medical Electrical Equipment - Part 2- Particular Requirements for Basic Safety and Essential Performance- Sec 49 Multifunction Patient Monitors IS/IEC 80601 (Part 2 / Sec 49): 2018

विभिन्न उत्पादों के लिए भारतीय मानक ब्यूरो) अनुरूपता मूल्यांकन (विनियम, 2018 की योजना -I के तहत प्रमाणन के संचालन में एकरूपता और पारदर्शिता के लिए इस उत्पाद मैनुअल का उपयोग सभी क्षेत्रीय / शाखा कार्यालयों और लाइसेंसधारियों द्वारा संदर्भ सामग्री के रूप में किया जाएगा। दस्तावेज़ का उपयोग बीआईएस प्रमाणन प्राप्त करने के इच्छुक संभावित आवेदकों द्वारा भी किया जा सकता है।

This Product Manual shall be used as reference material by all Regional/Branch Offices & licensees to ensure uniformity of practice and transparency in operation of certification under Scheme-I of Bureau of Indian Standards (Conformity Assessment) Regulations, 2018 for various products. The document may also be used by prospective applicants desirous of obtaining BIS certification.

1.	मानक संख्या IS No.	:	IS 80601 (Part 2/ Sec 49): 2018
	शीर्षक Title	:	Medical Electrical Equipment - Part 2- Particular Requirements for : Basic Safety and Essential Performance- Sec 49 Multifunction Patient Monitors
	संशोधनों की संख्या No. of amendments	:	Nil
2.	नमूना दिशानिर्देश Sampling Guidelines		
a)	कच्चा माल Raw material	:	Components – As per Cl 4.8 of IS 13450 (Part 1):2018
b)	समूहीकरण दिशानिर्देश Grouping Guidelines	:	Please refer ANNEX – A
c)	नमूने का परिमाण Sample Quantity	:	1 machine (with all the accessories)
3.	परीक्षण उपकरणों की सूची List of Test Equipment	:	Please refer ANNEX – B
4.	निरीक्षण और परीक्षण की स्कीम Scheme of Inspection and Testing	:	Please refer ANNEX – C

5.	एक दिन में संभावित परीक्षण Possible tests in a day	:	• Display of PATIENT data, CI 201.15.101 • Units of measurement, CI 201.15.102 • Indicator of operation, CI 201.15.4.4.101 • Determination of ALARM CONDITIONS and assignment of priority, CI 208.6.1.2 • Indication of validity of measured values, CI 208.6.3.2.101 • Alarm Presets, CI 208.6.5 • Alarm Signal Inactivation States, CI 208.6.8 • Non-Latching and Latching Alarm Signals, CI 208.6.10 • Alarm System logging, CI 208.6.12 • Protection against electrical hazards from ME Equipment, CI 201.8 & CI 201.8.5.2.3 • Primary Operating Functions, CI 206.101 • The currents from, to or between Patient connections, CI 8.4.2(a) • The Leakage Currents from, to or between Accessible parts, CI 8.4.2(b) • Interruption of power supply, CI 201.11.8
6.	लाइसेंस का दायरा /Scope of the Licence:		
“License is granted to use Standard Mark as per <i>IS 80601 (Part 2/Sec 49):2018</i> with the following scope:			
Name of the product		Medical Electrical Equipment - Part 2- Particular Requirements for Basic Safety and Essential Performance- Sec 49 Multifunction Patient Monitors	
Model			
Type		Stationary (or) Fixed / Mobile / Portable / Hand-held	
Rating		Rated VoltageV (or) Rated Voltage Ranges(s)V, ac Frequency Hz, Single phase / Three Phase, Rated Current A, Input powerW	
Protection against electric shock		Class – I / Class – II Type BF Applied Parts / Type CF Applied Parts Defibrillation-Proof Applied Parts / Non-Defibrillation-Proof Applied Parts	
Mode of Operation		Continuous	
Protection against harmful ingress of water or particulate matter			
Suitability for use in oxygen rich environment		Suitable/ Not suitable	
Method of Sterilization		By ethylene oxide gas / By irradiation such as gamma ray / By moist heat such as by autoclave / By other methods validated and described by the manufacturer	
Physiological Parameters Monitored (displayable on ME)		Electrocardiography / Non-invasive blood pressure / Invasive blood pressure / Pulse oximetry / Temperature / Electroencephalography / Transcutaneous gas analysis / Respiratory gas analysis / Any other parameter to be declared by manufacturer	

Alarm System	Alarm system included / Component of a distributed alarm system
Display of Information	Displays information / Distributes the information for remote display
Accessories (to be used with the Medical Equipment)	

ANNEX -A

Grouping Guidelines

1. Manufacturer shall submit declaration for each of the parameters of each model as specified in scope of the licence (as per Sl. No. 6 above) of 'Multifunction Patient Monitors' they intend to cover in the scope of licence. Further, the manufacturer shall also declare the components and essential accessories (if any) supplied along with the model.
2. For Grant of Licence/ Change in Scope of Licence, each 'Model' of 'Multifunction Patient Monitors' is to be tested.
3. During the operation of the Licence, BO shall ensure that all the model(s) covered in the Scope of the Licence are tested in rotation to the extent possible.

ANNEX- B**List of Test Equipment & Documents
Major test equipment required to test as per the Indian Standard**

Sl. No.	Tests used in with Clause Reference	Test Equipment
1.	Power Input, 4.11	Power Input Test Apparatus/Power Analyzer / Multimeter
2.	Accessible Parts 5.9.2	Test Finger with accessibility test apparatus
3.	Humidity preconditioning treatment, 5.7	Conditioning chamber with temperature and humidity chamber
4.	Durability of markings, 7.1.3	Distilled water, ethanol 96%, Isopropyl alcohol, Stop watch
5.	Earth Leakage Current, 8.7.4.5, Touch Current, 8.7.4.6, Patient Leakage Current, 8.7.4.7	Figure 13 to Figure 18, Leakage Current Tester, metal foil, Test Pin as per Figure 8 or a particular tool as specified by manufacturer
6.	Impedance and current carrying capability, 8.6.4	AC/DC voltage source, Multimeter/Earth Resistance Tester
7.	Dielectric Strength, 8.8.3	High Voltage Tester
8.	Creepage Distances And Air Clearances 8.9	Vernier Caliper, Filler Gauge
9.	Patient leads or patient cables, Cl 201.8.5.2.3 of IS 80601-2-49	Simulator

Note 1: The above list is indicative only and may not be treated as exhaustive.

Note 2: The licensee shall maintain a Technical file and Risk Management File as per the provisions of IS/IEC 80601 (Part 2/ Sec 49): 2018 for each model in the scope of license.

Note 3: The Technical File and Risk Management File shall be maintained by the licensee till the End of Life for each model.

ANNEX C

Scheme of Inspection and Testing

1. ENSURING COMPLIANCE THROUGH TESTING- It is expected that manufacturers (licensees/applicants) will establish a suitably equipped and staffed in house laboratory (In house testing facility) for testing at least those parameters of the Indian Standard which require routine testing for ensuring quality of the product. This includes in-process controls as may be defined and put in place by the manufacturer and testing parameters/requirements which can only be performed in the factory.

1.1 For MSME manufacturers, the requirement of maintaining a laboratory/in-house testing facility is optional. 1.2 For the guidance of manufacturers, Table 1 giving the recommended levels of control is given below. Column 2 of Table 1 indicates where test equipment is required in house i.e “R” or whether it can be subcontracted as “S”. Subcontracting is permitted to BIS recognized/empanelled laboratory or any other laboratory having valid accreditation as per IS/ISO/IEC 17025, test reports issued by the laboratories shall be available for inspection by BIS.

1.3 Large Scale manufacturers (including foreign manufacturers) shall maintain an in-house laboratory, which shall be suitably equipped (as given in column 2 of Table 1) and staffed, where different tests given in the specification shall be carried out in accordance with the method given in the specification. They shall also implement a calibration plan for the test equipment.

Alternatively, in lieu of an in-house laboratory, these manufacturers can also utilize the provisions of Sharing of testing facilities as per the Annexure-X (Relaxation in test facilities) of the guidelines for Grant of Licence available on BIS website www.bis.gov.in. (Under Conformity Assessment>Product Certification Process)

1.4 Further, MSME manufacturers may utilize common cluster based facilities as per guidelines for the utilization of cluster based test facilities by MSMEs or the provisions of Sharing of testing facilities as per 1.2 above (guidelines available on BIS website www.bis.gov.in (Under Conformity Assessment>Product Certification Process)).

2. QUALITY ASSURANCE PLAN

2.1 It is expected that manufacturers (licensees/applicants) will implement a Quality Control Assurance Plan i.e. a plan of regular testing and in-process controls, designed to ensure that the product bearing the Standard Mark conforms to all requirements of the Indian Standard.

2.2 The manufacturers shall define a Quality Control system defining the control unit (i.e. lot/batch etc.) and the levels of control (i.e. the frequency and number of samples for conducting the different tests as per the Indian Standard) and submit the same to BIS Branch Office for information. The manufacturer shall comply with the same and maintain test records in accordance with para 2.5.

2.3 RECOMMENDED LEVELS OF CONTROL/CONTROL UNIT:

2.3.1 As mentioned above, recommended levels of control are given in **Table 1 for guidance only.**

2.3.2 The tests as indicated in Table 1 and at the levels of control in column 3 of Table 1 or the Quality Assurance Plan as submitted by the manufacturer to BIS may be carried out on the whole production of the factory which is covered by this plan and records maintained in accordance with para 2.5.

2.4 However, all manufacturers shall ensure compliance of their products to all the requirements of the Indian Standard.

2.5 TEST RECORDS- The manufacturers maintaining an in-house laboratory or utilizing common cluster based facilities or shared test facilities shall maintain test records for the tests carried out to establish conformity. For the

tests being subcontracted to BIS recognized/empanelled laboratory or any other laboratory having valid accreditation as per IS/ISO/IEC 17025, test reports issued by the laboratories shall be available for inspection by BIS.

2.5.1 TECHNICAL FILE- Licencee shall maintain a Technical File for each model being manufactured by them. The Technical file shall be submitted to BIS at the time of Grant of Licence/ inclusion/ surveillance.

2.5.1.1 The Technical File shall be maintained till the End of Life of the model.

2.5.1.2 Licencee shall declare the frequency for periodic review in Technical file to BIS and record of review to be maintained. In case of any change in the Technical file due to any periodic review/ change in the product, the licensee shall inform the same to BIS at the earliest and provide the copy of the latest version of Technical File to BIS within 30 days of issue of the revised/ modified Technical File.

2.5.1.3 For changes in the product which shall affect the Scope of the Licence, the modified product shall be treated as a different model and separate Technical File shall be maintained for the same.

2.5.1.4 The technical file shall consist of the following documents:

- A general description of the product;
- Product Manual and/ or Instructions for use;
- Instructions for repair/ service for the product/ components, if required;
- Frequency of periodic repair/ service for the product/ components, if required;
- Conceptual product design and manufacturing drawings and, where appropriate, list of components, electrical circuit design, etc with descriptions and explanations needed for the understanding of these drawings;
- A list of the all Indian/ International standards applied in full or in part during the conformity assessment process with copy of test certificate/ test report;
- Test reports/ Test Certificates and risk assessment file for the product and/ or components as per relevant Clauses of the Indian Standard;
- Copies of conformity assessment documentation for critical product components;
- Copy of BIS licences for the product/ components, wherever applicable;
- A copy of the Declaration of Conformity for other International Standards;
- Other documents, if found necessary by the manufacturer and/ or by BIS may also be a part of the Technical File

2.5.2 RISK MANAGEMENT FILE- The licensee shall maintain a Risk Management File for each model in the scope of licence as per the requirement of IS / IEC 80601 (Part 2/ Sec 49): 2018. The Risk Management File shall be submitted to BIS at the time of Grant of Licence/Inclusion/Surveillance.

2.5.2.1 The Risk Management File shall be maintained till the End of Life of each model.

2.5.2.2 Licencee shall declare the frequency for periodic review in Risk Management File to BIS and record of review in the file to be maintained. In case of any change in the Risk Management File due to any periodic review/ change in the product, the licensee shall inform the same to BIS at the earliest and provide the copy of the latest version of Risk Management File to BIS within 30 days of issue of the revised/ modified Technical File.

3. PACKING AND MARKING - The Standard Mark as given in the Schedule of the licence shall be incorporated legibly and indelibly on each Multifunction Patient Monitors provided always that the material so marked conforms to each requirement of the specification.

3.1 Packing and Marking shall be done as per Cl. 201.7 of IS / IEC 80601 (Part 2/ Sec 49): 2018.

3.2 Additional Marking requirements: The material shall also be marked with the following additional

requirement on each product and packaging:

- a) “For BIS certification details please visit www.bis.gov.in”

3.3 All the production which conforms to the Indian Standard and covered under the scope of this licence shall be marked with the Standard Mark.

4. HYGIENIC CONDITIONS (if applicable) – NA.

5. REJECTION - Disposal of non-conforming product shall be done in such a way so as to ensure that there is no violation of provisions of BIS Act,2016.

TABLE-1

(1)				(2)	(3)		
Test Details				Test Equipment requirement R: Required (or) S: Sub-contracting permitted	Levels of Control		
Cl.	Requirement	Test Method			No. of Sample	Frequency	Remarks
		Clause	Reference				
201.4 General requirements							
201.4.1	Conditions for application to ME EQUIPMENT or ME SYSTEMS	4.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	
201.4.2	Risk management process for ME equipment or ME systems	4.2	IS 13450 (Part 1)	R	One	Each Medical Equipment	
201.4.3	Essential Performance						
201.4.3.101	Additional ESSENTIAL PERFORMANCE requirements	201.4.3.101	IS 13450 (Part 1) IS/IEC 80601-2-49	R	One	Each Medical Equipment	
201.4.4	EXPECTED SERVICE LIFE	201.4.4	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
201.4.5	Alternative RISK CONTROL measures or test methods for ME EQUIPMENT or ME SYSTEMS	201.4.5	IS 13450 (Part 1) IS/IEC 80601-2-49	R			

201.4.6	ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT	201.4.6	IS 13450 (Part 1) IS/IEC 80601-2-49	R			
201.4.7	SINGLE FAULT CONDITION for ME EQUIPMENT	13.1 13.2	IS 13450 (Part 1) IS/IEC 80601-2-49	S			
201.4.8	Components of ME EQUIPMENT	201.4.8	IS 13450 (Part 1) IS/IEC 80601-2-49	R	One	Each Consignm ent	No further testing if accompanied with Test Certificate or ISI Mark or covered in Risk Management file
201.4.9	Use of COMPONENTS WITH HIGH-INTEGRITY CHARACTERISTICS in ME EQUIPMENT	201.4.9	IS 13450 (Part 1) IS/IEC 80601-2-49				
4.10	Power supply						
4.10.1	Source of power for ME EQUIPMENT	4.10.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
4.10.2	SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS	4.10.2	IS 13450 (Part 1)	R			
4.11	Power input	4.11	IS 13450 (Part 1)	R			
201.7	ME Equipment identification, marking and documents						
201.7	ME Equipment Identification, marking and documents	7	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
201.7.2	Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts	7.2 201.7.2	IS 13450 (Part 1) IS/IEC 80601-2-49	R			

201.7.3	Marking on the inside of ME EQUIPMENT or ME EQUIPMENT parts	7.3	IS 13450 (Part 1)	R			
201.7.4	Marking of controls and instruments	7.4 201.7.4.3	IS 13450 (Part 1) IS/IEC 80601-2-49	R			
201.7.5	Safety signs	7.5		R			
201.7.6	Symbols	7.6	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
201.7.7	Colours of the insulation of conductors	7.7		R			
201.7.8	Indicator lights and controls	7.8		R			
201.7.9	ACCOMPANYING DOCUMENTS	7.9 201.7.9.2.2 201.7.9.2.9		R			
201.8	Protection against electrical hazards from ME Equipment						
8.1	Fundamental rule of protection against electric shock	8.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.2	Requirements related to power sources	8.2	IS 13450 (Part 1)	R			
8.3	Classification of APPLIED PARTS	8.3 201.8.3	IS 13450 (Part 1) IS/IEC 80601-2-49	R			

8.4	Limitations of voltage, current or energy						
8.4.2 (a)	The currents from, to or between Patient connections	8.7.4	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.4.2 (b)	The Leakage Currents from, to or between Accessible parts	8.7.4	IS 13450 (Part 1)	R	One		
8.4.2 (c)	Other accessible parts	8.4.2	IS 13450 (Part 1)	R	One		
8.4.2 (d)	Internal parts that are accessible	8.4.2	IS 13450 (Part 1)	R	One		
8.4.3	ME equipment intended to be connected to a power source by a plug	8.4.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.4.4	Internal capacitive circuits	8.4.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.5	Separation of Parts						
8.5.1.2	Means of Patient Protection (MOPP)	8.5.1.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.5.1.3	MEANS OF OPERATOR PROTECTION (MOOP)	8.5.1.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.5.2	Separation of PATIENT CONNECTIONS	8.5.2 201.8.5.2.3	IS 13450 (Part 1) IS/IEC 80601-2-49	S	One	Once in 03 years	Refer Note 1
8.5.3	MAXIMUM MAINS VOLTAGE	8.5.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.5.4	WORKING VOLTAGE	8.5.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1

8.5.5	DEFIBRILLATION - PROOF APPLIED PARTS	8.5.5	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.5.5.1	Defibrillation Protection	8.5.5.1 201.8.5.5.1	IS 13450 (Part 1) IS/IEC 80601-2- 49	S	One	Once in 03 years	Refer Note 1
8.5.5.2	Energy Reduction Test	8.5.5.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.6	Protective earthing, functional earthing and potential equalization of ME EQUIPMENT						
8.6.1	Applicability of requirements	8.6.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.6.2	PROTECTIVE EARTH TERMINAL	8.6.2	IS 13450 (Part 1)	R			
8.6.3	Protective earthing of moving parts	8.6.3	IS 13450 (Part 1)	R			
8.6.4	Impedance and current- carrying capability	8.6.4	IS 13450 (Part 1)	R			
8.6.5	Surface Coatings	8.6.5	IS 13450 (Part 1)	R			
8.6.6	Plugs and sockets	8.6.6	IS 13450 (Part 1)	R			
8.6.7	POTENTIAL EQUALIZATION CONDUCTOR	8.6.7	IS 13450 (Part 1)	R			
8.6.8	FUNCTIONAL EARTH TERMINAL	8.6.8	IS 13450 (Part 1)	R			
8.6.9	CLASS II ME EQUIPMENT	8.6.9	IS 13450 (Part 1)	R			
8.7	LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS						

8.7.1	General Requirements	8.7.1	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.7.2	SINGLE FAULT CONDITIONS	8.7.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.7.3	Allowable values	8.7.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.7.4	Measurements	8.7.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.8	Insulation						
8.8.2	Distance through solid insulation or use of thin sheet material	8.8.2	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.8.3	Dielectric strength	8.8.2	IS 13450 (Part 1)	R			
8.8.4	Insulation other than wire insulation						
8.8.4.1	Mechanical strength and resistance to heat	8.8.4.1	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.8.4.2	Resistance to environmental stress	8.8.4.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.9	CREEPAGE DISTANCES and AIR CLEARANCES						
8.9.1	Values	8.9.1	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.9.2	Applications	8.9.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.9.3	Spaces filled by insulating compound	8.9.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1
8.9.4	Measurement of CREEPAGE	8.9.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer Note 1

14

	DISTANCES AND AIR CLEARANCES						
8.10	Components and wiring						
8.10.1	Fixing of components	8.10.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.10.2	Fixing of wiring	8.10.2	IS 13450 (Part 1)	R			
8.10.3	Connections between different parts of ME EQUIPMENT	8.10.3	IS 13450 (Part 1)	R			
8.10.4	Cord-connected HAND-HELD parts and cord- connected foot-operated control devices	8.10.4	IS 13450 (Part 1)	R			
8.10.5	Mechanical protection of wiring	8.10.5	IS 13450 (Part 1)	R			
8.10.6	Guiding rollers for insulated conductors	8.10.6	IS 13450 (Part 1)	R			
8.10.7	Insulation of internal wiring	8.10.7	IS 13450 (Part 1)	R			
8.11	MAINS PARTS, components and layout						
8.11.1	Isolation from Supply mains	8.11.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.11.2	Multiple Socket Outlets	8.11.2	IS 13450 (Part 1)	R			
8.11.3	Power Supply Cords						

8.11.3.1	Application	8.11.3.1	IS 13450 (Part 1)	R	One		Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
8.11.3.2	Types	8.11.3.2	IS 13450 (Part 1)	R				
8.11.3.3	Cross-sectional area of POWER SUPPLY CORD conductors	8.11.3.3	IS 13450 (Part 1)	R				
8.11.3.4	Appliance Couplers Cord anchorage Cord guards	8.11.3.4 8.11.3.5 8.11.3.6	IS 13450 (Part 1) or IS/IEC 60320-1	R				
8.11.4	Mains terminal devices	8.11.4	IS 13450 (Part 1)	R				
8.11.5	Mains fuses and over current releases	8.11.5	IS 13450 (Part 1)	R				
8.11.6	Internal wiring of the Mains Part	8.11.6	IS 13450 (Part 1)	R				
201.9	Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS							
9.1	MECHANICAL HAZARDS of ME EQUIPMENT	9.1	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1	
9.2	MECHANICAL HAZARDS associated with moving parts	9.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1	
9.3	MECHANICAL HAZARD associated with surfaces, corners and edges	9.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1	
9.4	Instability HAZARDS	9.4,	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1	

9.5	Expelled parts HAZARD	9.5	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
9.6	Audible acoustic energy	9.6	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
9.7	Pressure vessels and parts subject to pneumatic and hydraulic pressure	9.7	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
9.8	MECHANICAL HAZARDS associated with support systems	9.8	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
201.10	Protection against unwanted and excessive radiation HAZARDS						
10.1	X-Radiation	10.1	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
10.2	Alpha, beta, gamma, neutron and other particle radiation	10.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
10.3	Microwave radiation	10.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
10.4	Lasers	10.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
10.5	Other visible electromagnetic radiation	10.5	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
10.6	Infrared radiation	10.6	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
10.7	Ultraviolet radiation	10.7	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
11	Protection against excessive temperatures and other HAZARDS						

11.1	Excessive temperatures in ME EQUIPMENT	11.1	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
11.2	Fire Prevention	11.2	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
11.3	Constructional requirements for fire ENCLOSURES of ME EQUIPMENT	11.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
11.4	ME EQUIPMENT and ME SYSTEMS intended for use with flammable anaesthetics	11.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
11.5	ME EQUIPMENT and ME SYSTEMS intended for use in conjunction with flammable agents	11.5	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
11.6	Overflow, spillage, leakage, ingress of water or particulate matter, cleaning, disinfection, sterilization and compatibility with substances used with the ME EQUIPMENT	11.6, 201.11.6.5	IS 13450 (Part 1) IS/IEC 80601-2-49	R	One	Every Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
11.7	Biocompatibility of ME EQUIPMENT and ME SYSTEMS	11.7	IS 13450 (Part 1)	R	One	Once in 03 years	Refer note 1
11.8	Interruption of the power supply / SUPPLY MAINS to ME EQUIPMENT	201.11.8 201.11.8.101	IS/IEC 80601-2-49	R	One	Every Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
12	Accuracy of controls and instruments and protection against hazardous outputs						

12.1	Accuracy of controls and instruments	12.1 201.12.1	IS 13450 (Part 1)	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
12.2	USABILITY of ME EQUIPMENT	12.2	IS 13450 (Part 1)	R			
12.3	ALARM SYSTEMS	201.12.3	IS/IEC 80601-2-49	R			
12.4	Protection against hazardous output	12.4.1 12.4.2 12.4.3 12.4.4	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
13	HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT						
13	HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	13	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)						
14	Programmable Electrical Medical Systems (PEMS)	14	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
15	Construction of ME EQUIPMENT						
15.1	Arrangements of controls and indicators of ME EQUIPMENT	15.1 201.15.101 201.15.102	IS 13450 (Part 1) IS/IEC 80601-2-49	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
15.2	Serviceability	15.2	IS 13450 (Part 1)	R			
15.3	Mechanical strength	15.3	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1

15.4	Indicators	15.4 201.15.4.4	IS 13450 (Part 1) IS/IEC 80601-2-49	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
16	ME SYSTEMS						
16	ME SYSTEMS	16	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS						
17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	17	IS 13450 (Part 1)	S	One	Once in 03 years	Refer note 1
202	Electromagnetic disturbances – Requirements and tests						
202	Electromagnetic disturbances – Requirements and tests	202	IS 13450-1-2 IS/IEC 80601-2-49	S	One	Once in 03 years	Refer note 1
202.7	Electromagnetic EMISSIONS requirements for ME EQUIPMENT and ME SYSTEMS	202.7	IS 13450-1-2 IS/IEC 80601-2-49	S	One	Once in 03 years	Refer note 1
202.8.1	IMMUNITY TEST LEVELS	202.8.1 202.8.101 202.8.102	IS 13450-1-2 IS/IEC 80601-2-49	S	One	Once in 03 years	Refer note 1
206	Usability	206	IS 13450-1-6	R	One	Each Medical Equipment	Testing and assessment of the product to be done as specified in the Technical File and Risk Assessment File
208	General requirements, tests and guidance for ALARM SYSTEMS IN MEDICAL ELECTRICAL	208 208.6.1.2 208.6.3.2 208.6.5 208.6.8	IS 13450-1-8 IS/IEC 80601-2-49	R			

	EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	208.6.10 208.6.11.1 208.6.12					
--	---	------------------------------------	--	--	--	--	--

Note-1: In addition, the test(s) shall be done whenever there is a change in design of the product, components or parts.

Note-2: Report of the Factory Sample drawn for complete testing by BIS will be shared with the manufacturer. During the operation of licence, licensee has to ensure that all the model(s) covered in the scope of the licence are tested in rotation.